

Non-destructive Testing - Qualification of radiographic film digitisation systems — Part 2: Minimum Requirements

Contents

Introduction	1
1 Scope.....	1
2 Normative references.....	1
3 Terms and definitions.....	2
4 Digitisation quality classes.....	3
5 Minimum requirements for film digitisation quality classes.....	3

Foreword

This document has been prepared by Technical Committee CEN/TC 138, "Non-destructive testing", the secretariat of which is held by AFNOR.

This document is currently submitted to the Formal Vote.

EN 14096 comprises a series of European Standards for radiographic film digitisation systems which is made up the following:

- | | |
|------------|---|
| EN 14096-1 | Non-destructive testing — Qualification of radiographic film digitisation systems — Part 1: Definitions, quantitative measurements of image quality parameters, standard reference film and qualitative control |
| EN 14096-2 | Non-destructive testing — Qualification of radiographic film digitisation systems — Part 2: Minimum requirements |

Introduction

Radiographic film systems are used for industrial inspection by X and gamma rays. To apply modern means of computer support for analysis, transmission and storage the information stored in the radiographic film has to be converted into digital data (digitisation). This standard defines minimum requirements to ensure that the relevant information for evaluation of the digital data is preserved during the film digitisation process.

1 Scope

This European Standard specifies three film-digitisation quality classes for the requirements of non-destructive testing. The selected class depends on the radiation energy, penetrated material thickness and the quality level of the original radiographic film. This standard does not address signal processing, display and storage of the digitised data.

2 Normative references

This **Fehler! Verweisquelle konnte nicht gefunden werden.** incorporates by dated or undated reference, provisions from other publications. These normative references are cited at the appropriate places in the text and the publications are listed hereafter. For dated references, subsequent amendments to or revisions of any of these publications apply to this **Fehler! Verweisquelle konnte nicht gefunden werden.** only when incorporated in it by amendment or revision. For undated references the latest edition of the publication referred to applies (including amendments).

EN 444, *General principles for the radiographic examination of metallic materials using X-rays and gamma-rays.*

EN 1435, *Non-destructive examination of welds — Radiographic examination of welded joints.*

prEN 12681, *Founding — Radiographic inspection.*

prEN 14096-1, *Non-destructive testing — Qualification of radiographic film digitisation systems — Part 1: Definitions, quantitative measurements of image quality parameters, standard reference film and qualitative control.*

ISO 5579, *Non-destructive testing — Radiographic examination of metallic materials by X- and gamma rays — Basic rules.*

3 Terms and definitions

For the purposes of this **Fehler! Verweisquelle konnte nicht gefunden werden.**, the following terms and definitions apply.

3.1 radiographic film digitisation system (Digitiser)

sequential application of the two functions below:

- a) Detection of the diffuse transmittance of a small unit area of the film (Pixel, picture element) by means of an optical detector, giving an electric output signal (geometrical digitisation);
- b) conversion of the above electrical signal into a numerical value (densitometrical digitisation)

3.2 pixel size P

geometrical centre-to-centre distance between adjacent pixels in a row (horizontal pitch) or column (vertical pitch) of the scanned image

3.3 optical density D

logarithmic value to the base 10 of the diffuse light intensity ratio in front of (I_0) and behind (I_D) the radiographic film according to equation 1:

$$D = \lg \frac{I_0}{I_D} \quad (1)$$

3.4 spatial frequency f

described by a sinusoidal intensity variation along a geometrical axis

The period of this function is measured in number of line pairs per millimetre (lp/mm).

3.5 modulation transfer function MTF

normalised magnitude of the Fourier-transform (FT) of the differentiated spatial optical density edge spread function (ESF) (see prEN 14096-1, figure 1)

It describes the unsharpness function of the digitiser (contrast transmission as a function of the object size).

NOTE This MTF calculation is based on optical densities, which correspond to the X-ray dose.

3.6 digital resolution in bit

number of bits provided by the analogue-to-digital converter of the digitiser used for densitometrical digitisation

NOTE A digital resolution of N bits corresponds to 2^N digital values.

3.7 density contrast sensitivity DD_{CS}

minimum density variation of the film, which is resolved by the digitiser

This is mostly determined by the digitisation noise of the digitise (quantum noise of the light detector).

NOTE The determination of this value is described in part 1, clause 4.1.5.

3.8 density range D_R

range of maximum and minimum optical densities, which can be measured by the digitiser

Depending on the construction of the digitiser, this density range can be split into several working ranges (e.g. by a different illumination power and/or a different detector integration time).

3.9 working range D_{WR}

range of optical densities, where the digitiser guarantees a minimum density contrast sensitivity in one single acquisition

Only in this working range the digitised data can be used for evaluation. Depending on the digitiser construction there can be more than one working range, e.g. for brighter or darker films.

4 Digitisation quality classes

All radiographic film digitisation systems are subdivided into 3 quality classes DS, DB and DA :

DS - the enhanced technique which performs the digitisation with an insignificant reduction of signal-to-noise-ratio and spatial resolution.

Application field: digital archiving of films (digital storage).

DB - the enhanced technique, which permits some reduction of image quality.

Application field: digital analysis of films, original radiographic films have to be archived.

DA - the basic technique, which permits some reduction of image quality and further reduced spatial resolution.

Application field: digital analysis of films, original radiographic films have to be archived.

Each radiographic film digitisation system for NDT applications shall be identified with all working ranges of optical densities. It shall be classified by the quality class corresponding to table 1 and the maximum MTF 20 % value, corresponding to prEN 14096-1, which can be performed by this system.

EXAMPLE A digitisation system of class DS 4,2 (scanner class DS (see table 1), maximum MTF 20 % = 4,2 lp/mm (see table 2)) can be applied for archiving of radiographs taken with X-rays above 200 keV or γ -rays and can be applied for all class DB and DA digitisation tasks.

5 Minimum requirements for film digitisation quality classes

5.1 Density range and working ranges of the film digitisation system

Table 1 defines the minimum density range of the radiographic digitisation system. In this density range the digitiser shall provide a density contrast sensitivity ΔD_{CS} with $\Delta D_{CS} \leq 0,02$. Depending on the construction of the digitiser, this density range can be split into several working ranges.

The minimum digital resolution is given for devices converting the digital value proportional to the optical density. If the digital value is converted proportional to the light intensity, the digital resolution has to be increased at least by 2 additional bits.

Table 1 — Minimum density range of the radiographic digitisation system with a minimum density contrast sensitivity

Parameter	Class DS	Class DB	Class DA
Density range ^a D_R	0,5 to 4,5	0,5 to 4,0	0,5 to 3,5
Digital resolution in bit	≥ 12	≥ 10	≥ 10
Density contrast sensitivity ΔD_{CS} within D_R	$\leq 0,02$	$\leq 0,02$	$\leq 0,02$
^a This density range may be splitted into separated working ranges.			

5.2 Minimum spatial resolution of film digitisation systems

Due to the energy dependence of the inherent unsharpness of industrial X-ray film systems the following parameters (table 2) shall be observed:

Table 2 — Minimum spatial resolution of film digitisation systems

Energy keV	Class DS		Class DB		Class DA	
	Pixel size μm	MTF 20 % lp/mm	Pixel size μm	MTF 20 % lp/mm	Pixel size μm	MTF 20 % lp/mm
≤ 100	15	16,7	50	5	70	3,6
> 100 to 200	30	8,3	70	3,6	85	3
> 200 to 450, Se-75, Yb-169	60	4,2	85	3	100	2,5
Ir-192	100	2,5	125	2	150	1,7
Co-60, > 1 MeV	200	1,25	250	1	250	1

NOTE 1 For the normal check corresponding to prEN 14096-1, the MTF 20 % value may be determined by the converging spatial resolution targets.

NOTE 2 Due to possible aliasing, the converging spatial resolution targets may give less accurate values than MTF measurements.

NOTE 3 For energies lower than 70keV the spatial resolution of the radiographic film can be better than the scanner resolution required by the class DS 16,7. In this case the spatial resolution of the scanner should be adapted to the film resolution, or the original radiographic film should be archived.

5.3 Digitisation class depending on the radiographic testing class

The standards ISO 5579, EN 444, EN 1435 or prEN 12681 define two radiographic testing classes A and B. Radiographs taken according to these standards shall be digitised corresponding to table 3.

Table 3 — Minimum digitisation class depending on the radiographic testing classes A and B, if radiographs are taken on the basis of ISO 5579, EN 444, EN 1435 or prEN 12681

wall thickness [mm] steel	Class DS	Class DB	Class DA
< 5	B	A	—
≥ 5	B	B	A

NOTE For the detection of cracks and fine imperfections Class DB or better shall be used.

After digitisation process all required image quality indicators shall be visible in the digital image as on the original radiographic film.